

Client/Code

Cobble Hill Improve. District
 *A
 P.O. Box 27
 Cobble Hill, BC
 V0R 1L0

Date 09Mar21 3:47p No. W159817
 Source Well
 Type of Sample water
 No. of Samples 2

TEL: (250) 743-2861
 1979chid@gmail.com


Comments Arrival temp.: 7.6C
 Sampler: Alan Seal

SAMPLE	DATE	TIME	NH ₃ -N (ug/L)	E.C. (uS/cm)	TKN (mg/L)	NO ₃ -N (mg/L)	NO ₂ -N (ug/L)
1 Fisher Road Well	09Mar21	09:32a	ND	206	ND	3.05	ND
2 Holland Road Well	09Mar21	09:48a	ND	377	ND	3.59	ND
Lab Blank			ND	ND	ND	ND	ND
S ₀			0.254	0.300	0.012	0.160	0.300
REF. VALUE			10.0	147	1.00	10.0	10.0
STD ± 2SD			9.68 ± 0.780	142 ± 12.2	0.966 ± 0.080	9.82 ± 1.01	10.6 ± 0.760

SAMPLE	DATE	TIME	TN (mg/L)	ORG.N (mg/L)	Turbidity (NTU)
1 Fisher Road Well	09Mar21	09:32a	3.05	ND	0.180
2 Holland Road Well	09Mar21	09:48a	3.59	ND	0.210
Lab Blank			ND		ND
S ₀			0.030		0.015
REF. VALUE			1.00		5.00
STD ± 2SD			1.08 ± 0.077		4.88 ± 0.399

SD = standard deviation; REF VALUE = primary or secondary reference material
 STD = secondary standard calibrated to primary standard reference material
 S₀ = standard deviation at zero analyte concentration; method detection limit
 is generally considered to be 3x S₀ value
 ND = none detected n/a = not applicable

 R. Bilodeau
 Analytical Chemist



 H. Hartmann
 Sr. Analytical Chemist

